



OXYGEN CONCENTRATOR

- France imported CECA molecular sieve
- High oxygen purity 93%±3%
- Long life span more than 18000 hours
- Suitable for 24 hours continuously work
- 2 years warranty
- CE ISO approved
- OEM available

olv

SPECIFICATION

Model	OLV-5A
Purity Oxygen	≥93%±3 (V/V)
Voltage	AC220V±22V, 50Hz±1Hz/AC110V±15V, 60Hz±1Hz
Outlet Pressure	20kPa~60kPa
Flow Rate	0-5L/min
Sound Level	less than 45db
Power Consumption	320Watts
Net Weight	15kg
Dimension	34.4×30.6×56.5 (cm)
Warranty	2 years
Standard Accessories	1× Humidifier bottle 1× Primary filter 1× HEPA filter 2× Nasal oxygen cannula 1× User manual

CERTIFICATE

EC Certificate
 Production Quality Assurance System
 Directive 93/42/EEC on Medical Devices (MDD), Annex V
 (Devices in Class IIa, IIb or IIc)
 No. G2 103099 0002 Rev. 00

EC-Representative: MedPuh GmbH
 Mens-van-der-Ruhe-Strasse 6, 80807 München, GERMANY

Product Category(ies): Oxygen Concentrator

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for manufacture and final inspection of the respective devices / device categories in accordance with MDD Annex V. This quality assurance system conforms to the requirements of this Directive and is subject to periodic surveillance. For marking of class IIa and IIb devices an additional Annex III certificate is mandatory. See also notes overleaf!

Report No.: SH19164201
Valid from: 2019-04-23
Valid until: 2022-04-24

Date: 2019-04-25
S. Prütz
 Stefan Prütz

Page 1 of 2
 TÜV SÜD Product Service GmbH is a Notified Body with identification no. 0123
 TÜV SÜD Product Service GmbH • Certification Body • Fildersstraße 65 • 80339 Munich • Germany

Certificate
 No. Q6 103099 0001 Rev. 00

Certification Mark:

Scope of Certificate: Production and Distribution of Oxygen Concentrator

Applied Standard(s): EN ISO 13485:2016
 Medical devices - Quality management systems - Requirements for regulatory purposes
 (ISO 13485:2016)
 DIN EN ISO 13485:2016

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned above has established and is maintaining a quality management system (including substance 7.3), which meets the requirements of the listed standard(s). See also notes overleaf!

Report No.: SH19164201
Valid from: 2019-04-23
Valid until: 2022-04-22

Date: 2019-04-23
S. Prütz
 Stefan Prütz

Page 1 of 1
 TÜV SÜD Product Service GmbH • Certification Body • Fildersstraße 65 • 80339 Munich • Germany